

1000 Volt Cat III/Cat IV Fuse Probe Kit

Part Number:- 131350/R4

Features GS38 Compliant

1.2 Metre Double Skin Silicone Cable.

Bi-coloured layers.

1.0mm² c.s.a. 3.6mm Ø Nom.o/d Multistranded very flexible copper

core

Temp. Rating -10°C to +150°C Fuse Rating 500mA/1000V/50kA

Packed in silver wallet.

IEC 61010 Rating.

Measurement Category. 1000V CAT III/1000V CAT IV

Current Rating. As Per Fuse

Insultion Category. Double/Reinforced.

Pollution Degree. Two (2)

Silvertronic confirm that the product(s) listed have been designed, manufactured and tested to and are therefore fully compliant to:

IEC 61010-031:2015. (International CB Scheme). EN 61010-031:2015. (European). UL 61010-031, 2nd Edition. (United States). CSA C22.2 No 61010-031, 2nd Edition (Canada).

All products marked/labelled as per IEC requirements. UL and CSA labelling upon request.

All testing carried out by independent third party test house and CB Test Laboratory,

MET Laboratories Inc. Baltimore MD 21230 USA

The product(s) also meet the requirements of the Low Voltage Directive 2014/35/EU and where applicable comply with HSE GS.38.